




Congratulations on your purchase of this single use cold chain data logger. This data logger is specifically designed for monitoring the transportation of products subject to cold chain requirements. To avoid any damaging, please read through this instruction manual before using. The logger is calibrated before shipment.

OPERATING INSTRUCTION



EASY HANDLING

- 1 Press the **Start** button to enable Halveon T-Stream™ SU.
- 2 Insert enabled Halveon T-Stream™ SU in the delivery.
- 3 Remove the device from the delivery, press **Stop** to disable it. Connect the data logger directly to the USB Port of the computer.
- 4 Automatic report download (PDF & CSV format) without any software installation.



OPERATING INSTRUCTIONS

- 1 Push **Start/Stop** button for 5 seconds to start recording. Then **GREEN LED** flashes 5 times. This indicates a successful starting. Tear off serial number label and paste on required area.
- 2 Fast press button once at any time during the recording status. The led will flash 1 time to indicate the different status. See details in below tables.
- 3 Push **Start/Stop** button for 5 seconds to stop recording. Then **RED LED** flashes 5 times which indicates a successful stopping.
- 4 After stopping record, tear off the transparent plastic cover and plug into USB port of a computer. The PDF report will be generated automatically and can be found in the device folder.

OPERATING INDICATIONS

START/STOP

STATUS	ACTION	LED CONFIRMATION
START LOGGER	5 SEC PUSHING BUTTON	GREEN LED FLASHING 5 TIMES ●●●●
STOP LOGGER	5 SEC PUSHING BUTTON	RED LED FLASHING 5 TIMES ●●●●

STATUS REQUEST BY PUSHING BUTTON 1 TIME

STATUS	ACTION	LED CONFIRMATION
NOT STARTED	PUSH BUTTON 1 TIME	GREEN LED AND RED LED ARE FLASHING 1 TIME
RECORDING - OK	PUSH BUTTON 1 TIME	GREEN LED FLASHING 1 TIME ●
RECORDING - ALARM	PUSH BUTTON 1 TIME	RED LED FLASHING 1 TIME ●